

TLN

**Department of Forensic Science**  
**Breath Alcohol Section**  
**Instrument Location History**  
**1/1/1990 To 12/3/2014**

**Instrument Serial Number: 010583**

**As of 03-Dec-14**

<b>Start Date</b>	<b>End Date</b>	<b>Agency Name</b>
02-Dec-14		DFS Central Lab
25-Jun-14	02-Dec-14	Lancaster County Sheriff's Office
03-Jun-14	25-Jun-14	DFS Central Lab
03-Dec-08	03-Jun-14	Craig County SO
15-Aug-08	03-Dec-08	DFS Central Lab

TLN

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010583** Worksheet Start Date **12/2/2014**  
Location **Lancaster Co SO**  
Address **8293 Mary Ball Rd Lancaster VA 22503**  
DFS Technician **Timmy Neece, Jr** License No. **25387**  
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **771** Reference Barometer (mm HG) **773**  
Reference Barometer(RB)Serial # **009111** RB Calibration Due **5/19/2015**

## Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.311
0.300	0.304	0.294	0.313	Sample 2	0.310
Precision		sample min	sample max	Sample 3	0.310
0.001		0.310	0.311		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.103
0.100	0.101	0.097	0.104	Sample 2	0.103
Precision		sample min	sample max	Sample 3	0.103
0		0.103	0.103		

Dry gas standard Lot No. (with tank no.)


**AG332301-24**

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

Instrument Serial Number

010583

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

12/3/14

Issuing Analyst




Date

12/3/14

jpd

TLN




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>NEECE, JR., TIMMY, L</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>25387</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>12/02/2014</b>
TEST EQUIPMENT NUMBER <b>010583</b>		

RESULTS: TIME SAMPLE TAKEN 09:38 EST  
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 02 Dec 2014 at 10:08

MMB  
TLN

## Test Results

Instrument Serial Number 010583

Test # 000350 Subject Test

Test Location 1 Department of  
Test Date 02 Dec 2014

Test Location 2 Forensic Science  
Test Time 09:32  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name NEECE, JR.  
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

License Number 25387

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 02 Dec 2014 End Time 09:40

Result Time 09:38

Result Date 02 Dec 2014

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 09:32

Data Type BLK

Sample Value 0.000

Sample Time 09:33

Data Type CHK

Sample Value 0.103

Sample Time 09:34

Data Type BLK

Sample Value 0.000

Sample Time 09:35

Data Type SUBJ

Sample Value 0.000

Sample Time 09:36

Data Type BLK

Sample Value 0.000

Sample Time 09:37

Data Type SUBJ

Sample Value 0.000

Sample Time 09:38

Data Type BLK

Sample Value 0.000

Sample Time 09:39

Standard Type Dry Gas Std

Standard Value 0.101

Standard Lot Number AG332301-24

Standard Expiration Date 11/19/2015

Tank Pressure 977

Barometric Pressure 771 mmHg

Blow Sample Number 1 Blow Duration 2.87 sec

Blow Volume 1909 cc

End-of-Blow Time 09:36

Blow Sample Number 2 Blow Duration 2.98 sec

Blow Volume 1969 cc

End-of-Blow Time 09:38

Tamper Evident Stamp 21c153f5

Test Status Code 0

Test Status Success

TLN

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010583      Test Number: 351

Test Date: 12/02/2014      Test Time: 09:42 EST

Dry Gas Target: 0.101

Lot Number: AG332301-24      Exp Date: 11/19/2015

Tank Pressure: 973 psi      Barometric Pressure: 771 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:43
CHK	0.103	09:44
BLK	0.000	09:45
CHK	0.103	09:46
BLK	0.000	09:47
CHK	0.103	09:48

Test Status: *Success*

Calibration CRC: 6ED18318

TLN

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010583      Test Number: 352

Test Date: 12/02/2014      Test Time: 09:51 EST

Dry Gas Target: 0.304

Lot Number: AG428003-12      Exp Date: 10/07/2016

Tank Pressure: 531 psi      Barometric Pressure: 771 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:52
CHK	0.311	09:53
BLK	0.000	09:54
CHK	0.310	09:55
BLK	0.000	09:57
CHK	0.310	09:58

Test Status: *Success*

Calibration CRC: 6ED18318